

Manufacturer Disclosure Statement for Medical Device Security – MDS <sup>2</sup>			
DEVICE DESCRIPTION			
Device Category	Manufacturer	Document ID	Document Release Date
Clinical image viewer	Vital Images	VLC-10297	29-Jul-19
Device Model	Software Revision		Software Release Date
Vitrea View	7.5		29-Jul-19
Manufacturer or Representative Contact Information	Company Name	Manufacturer Contact Information	
	Representative Name/Position	5850 Opus Parkway, Suite 300, Minnetonka, MN 55343, USA (952) 487-9500	
<p><b>Intended use of device</b> in network-connected environment:                      Vitrea View software is a web-based, cross-platform, zero-footprint enterprise image viewer solution capable of displaying both DICOM and non-DICOM medical images.</p>			
MANAGEMENT OF PRIVATE DATA			
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.		Yes, No, N/A, or See	Note #
A	Can this <b>device</b> display, transmit, or maintain <b>private data</b> (including <b>electronic Protected Health Information [ePHI]</b> )?	Yes	1
B	Types of <b>private data</b> elements that can be maintained by the <b>device</b> :		
B.1	Demographic (e.g., name, address, location, unique identification number)?	Yes	—
B.2	Medical record (e.g., medical record #, account #, test or treatment date, <b>device</b> identification number)?	Yes	—
B.3	Diagnostic/therapeutic (e.g., photo/radiograph, test results, or physiologic data with identifying characteristics)?	Yes	—
B.4	Open, unstructured text entered by <b>device user/operator</b> ?	No	—
B.5	<b>Biometric data</b> ?	No	—
B.6	Personal financial information?	No	—
C	Maintaining <b>private data</b> - Can the <b>device</b> :		
C.1	Maintain <b>private data</b> temporarily in volatile memory (i.e., until cleared by power-off or	Yes	—
C.2	Store <b>private data</b> persistently on local media?	No	—
C.3	Import/export <b>private data</b> with other systems?	Yes	—
C.4	Maintain <b>private data</b> during power service interruptions?	No	—
D	Mechanisms used for the transmitting, importing/exporting of <b>private data</b> – Can the <b>device</b> :		
D.1	Display private data (e.g., video display, etc.)?	Yes	—
D.2	Generate hardcopy reports or images containing <b>private data</b> ?	No	—
D.3	Retrieve <b>private data</b> from or record <b>private data</b> to <b>removable media</b> (e.g., disk, DVD, CD-ROM, tape, CF/SD card, memory stick, etc.)?	Yes	—
D.4	Transmit/receive or import/export <b>private data</b> via dedicated cable connection (e.g., IEEE 1073, serial port, USB, FireWire, etc.)?	No	—
D.5	Transmit/receive <b>private data</b> via a wired network connection (e.g., LAN, WAN, VPN, intranet, Internet, etc.)?	Yes	—
D.6	Transmit/receive <b>private data</b> via an integrated wireless network connection (e.g., WiFi, Bluetooth, infrared, etc.)?	Yes	—
D.7	Import <b>private data</b> via scanning?	No	—
D.8	Other?	No	1
<p>Management of Private Data notes:                      1. Vitrea View relies on one of the following systems for image management:                      - Vitrea Data Stream (see VLC-09691 A Vitrea Data Stream 2.1 MDS2)                      - Vitrea Connection (see 2017.09.037 Vitrea Connection 7.2 MDS2)                      - Vitrea Image Management Systems (see VLC-09897 A Vitrea 7.8 MDS2)                      - 3rd party Vendor Neutral Archive</p>			

© Copyright 2013 by the National Electrical Manufacturers Association and  
the Healthcare Information and Management Systems Society.

HN 1-2013

Device Category <a href="#">Clinical image viewer</a>	Manufacturer <a href="#">Vital Images</a>	Document ID <a href="#">VLC-10297</a>	Document Release Date <a href="#">Jul-2019</a>
Device Model <a href="#">Vitrea View</a>	Software Revision <a href="#">7.5</a>	Software Release Date <a href="#">Jul-2019</a>	

**SECURITY CAPABILITIES**

Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.		Yes, No, N/A, or See	Note #
<b>1</b>	<b>AUTOMATIC LOGOFF (ALOF)</b> The <b>device's</b> ability to prevent access and misuse by unauthorized <b>users</b> if <b>device</b> is left idle for a period of time.		
1-1	Can the <b>device</b> be configured to force reauthorization of logged-in <b>user(s)</b> after a predetermined length of inactivity (e.g., auto-logoff, session lock, password protected screen cover)?	Yes	—
1-1.1	Is the length of inactivity time before auto-logoff/screen lock <b>user</b> or administrator configurable? (Indicate time [fixed or configurable range] in notes.)	Yes	1
1-1.2	Can auto-logoff/screen lock be manually invoked (e.g., via a shortcut key or proximity sensor, etc.) by the <b>user</b> ?	Yes	—
ALOF notes: <a href="#">1. Configurable</a>			
<b>2</b>	<b>AUDIT CONTROLS (AUDT)</b> The ability to reliably audit activity on the <b>device</b> .		
2-1	Can the <b>medical device</b> create an <b>audit trail</b> ?	Yes	1
2-2	Indicate which of the following events are recorded in the audit log:		
2-2.1	Login/logout	Yes	—
2-2.2	Display/presentation of data	Yes	—
2-2.3	Creation/modification/deletion of data	Yes	—
2-2.4	Import/export of data from <b>removable media</b>	Yes	—
2-2.5	Receipt/transmission of data from/to external (e.g., network) connection	Yes	—
2-2.5.1	<b>Remote service</b> activity	No	—
2-2.6	Other events? (describe in the notes section)	No	—
2-3	Indicate what information is used to identify individual events recorded in the audit log:		
2-3.1	<b>User ID</b>	Yes	—
2-3.2	Date/time	Yes	—
AUDT notes: <a href="#">1. The device can create audit logs that list user actions to ePHI data. However, the device does not create or delete ePHI.</a>			
<b>3</b>	<b>AUTHORIZATION (AUTH)</b> The ability of the device to determine the authorization of users.		
3-1	Can the <b>device</b> prevent access to unauthorized <b>users</b> through <b>user</b> login requirements or other mechanism?	Yes	—
3-2	Can <b>users</b> be assigned different privilege levels within an application based on 'roles' (e.g., guests, regular <b>users</b> , power <b>users</b> , administrators, etc.)?	Yes	—
3-3	Can the <b>device</b> owner/ <b>operator</b> obtain unrestricted administrative privileges (e.g., access operating system or application via local root or admin account)?	Yes	—
AUTH notes:			

Device Category	Manufacturer	Document ID	Document Release Date
Clinical image viewer	Vital Images	VLC-10297	Jul-2019
Device Model	Software Revision		Software Release Date
Vitrea View	7.5		Jul-2019
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			No, N/A, or See
<b>4 CONFIGURATION OF SECURITY FEATURES (CNFS)</b>			
The ability to configure/re-configure <b>device security capabilities</b> to meet <b>users'</b> needs.			
4-1	Can the <b>device</b> owner/operator reconfigure product <b>security capabilities</b> ?	Yes	—
CNFS notes:			
<b>5 CYBER SECURITY PRODUCT UPGRADES (CSUP)</b>			
The ability of on-site service staff, remote service staff, or authorized customer staff to install/upgrade <b>device's</b> security patches.			
5-1	Can relevant OS and <b>device</b> security patches be applied to the <b>device</b> as they become available?	Yes	1
5-1.1	Can security patches or other software be installed remotely?	Yes	1
CSUP notes:			
1. Vital Images, Inc. does not provide customers with Anti-Virus Software Definition Files, Operating System Security patches or Service Packs. Customer is responsible for managing hardware security and software update policies.			
<b>6 HEALTH DATA DE-IDENTIFICATION (DIDT)</b>			
The ability of the <b>device</b> to directly remove information that allows identification of a person.			
6-1	Does the <b>device</b> provide an integral capability to de-identify <b>private data</b> ?	No	—
DIDT notes:			
<b>7 DATA BACKUP AND DISASTER RECOVERY (DTBK)</b>			
The ability to recover after damage or destruction of <b>device</b> data, hardware, or software.			
7-1	Does the <b>device</b> have an integral data backup capability (i.e., backup to remote storage or <b>removable media</b> such as tape, disk)?	No	1
DTBK notes:			
1. Vital supports standard database and local disk backups, though the device does not retain any patient data.			
<b>8 EMERGENCY ACCESS (EMRG)</b>			
The ability of <b>device users</b> to access <b>private data</b> in case of an emergency situation that requires immediate access to stored <b>private data</b> .			
8-1	Does the <b>device</b> incorporate an <b>emergency access</b> ("break-glass") feature?	Yes	—
EMRG notes:			
<b>9 HEALTH DATA INTEGRITY AND AUTHENTICITY (IGAU)</b>			
How the <b>device</b> ensures that data processed by the <b>device</b> has not been altered or destroyed in an unauthorized manner and is from the originator.			
9-1	Does the <b>device</b> ensure the integrity of stored data with implicit or explicit error detection/correction technology?	N/A	—
IGAU notes:			

© Copyright 2013 by the National Electrical Manufacturers Association and  
the Healthcare Information and Management Systems Society.

HN 1-2013

Device Category	Manufacturer	Document ID	Document Release Date
Clinical image viewer	Vital Images	VLC-10297	Jul-2019
Device Model	Software Revision		Software Release Date
Vitrea View	7.5		Jul-2019
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			Yes, No, N/A, or See
			Note #
<b>10 MALWARE DETECTION/PROTECTION (MLDP)</b>			
The ability of the <b>device</b> to effectively prevent, detect and remove malicious software ( <b>malware</b> ).			
10-1	Does the <b>device</b> support the use of <b>anti-malware</b> software (or other <b>anti-malware</b>	Yes	1
10-1.1	Can the <b>user</b> independently re-configure <b>anti-malware</b> settings?	Yes	1
10-1.2	Does notification of <b>malware</b> detection occur in the <b>device user</b> interface?	No	1
10-1.3	Can only manufacturer-authorized persons repair systems when <b>malware</b> has been detected?	N/A	1
10-2	Can the device owner install or update <b>anti-virus software</b> ?	Yes	—
10-3	Can the device owner/ <b>operator</b> (technically/physically) update virus definitions on manufacturer-installed <b>anti-virus software</b> ?	N/A	1
MLDP notes:			
1. Vital Images, Inc. does not provide customers with Anti-Virus Software Definition Files, Operating System Security patches or Service Packs. Customer is responsible for managing hardware security and software update policies.			
<b>11 NODE AUTHENTICATION (NAUT)</b>			
The ability of the <b>device</b> to authenticate communication partners/nodes.			
11-1	Does the <b>device</b> provide/support any means of node authentication that assures both the sender and the recipient of data are known to each other and are authorized to receive transferred information?	No	—
NAUT notes:			
<b>12 PERSON AUTHENTICATION (PAUT)</b>			
Ability of the <b>device</b> to authenticate <b>users</b>			
12-1	Does the <b>device</b> support <b>user/operator</b> -specific username(s) and password(s) for at least one <b>user</b> ?	Yes	—
12-1.1	Does the device support unique <b>user/operator</b> -specific IDs and passwords for multiple users?	Yes	—
12-2	Can the <b>device</b> be configured to authenticate <b>users</b> through an external authentication service (e.g., MS Active Directory, NDS, LDAP, etc.)?	Yes	—
12-3	Can the <b>device</b> be configured to lock out a <b>user</b> after a certain number of unsuccessful logon attempts?	Yes	—
12-4	Can default passwords be changed at/prior to installation?	No	—
12-5	Are any shared <b>user</b> IDs used in this system?	Yes	—
12-6	Can the <b>device</b> be configured to enforce creation of <b>user</b> account passwords that meet established complexity rules?	Yes	—
12-7	Can the <b>device</b> be configured so that account passwords expire periodically?	Yes	—
PAUT notes:			
<b>13 PHYSICAL LOCKS (PLOK)</b>			
Physical locks can prevent unauthorized <b>users</b> with physical access to the <b>device</b> from compromising the integrity and confidentiality of <b>private data</b> stored on the <b>device</b> or on <b>removable media</b> .			
13-1	Are all <b>device</b> components maintaining <b>private data</b> (other than <b>removable media</b> ) physically secure (i.e., cannot remove without tools)?	N/A	—

PLOK

notes:

The device is a software component only and does not own or control the physical hardware.

© Copyright 2013 by the National Electrical Manufacturers Association and  
the Healthcare Information and Management Systems Society.

Device Category	Manufacturer	Document ID	Document Release Date
Clinical image viewer	Vital Images	VLC-10297	Jul-2019
Device Model	Software Revision		Software Release Date
Vitrea View	7.5		Jul-2019
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			No, N/A, or See
			Note #
<b>14</b>	<b>ROADMAP FOR THIRD PARTY COMPONENTS IN DEVICE LIFE CYCLE (RDMP)</b>		
	Manufacturer's plans for security support of 3rd party components within <b>device</b> life cycle.		
14-1	In the notes section, list the provided or required (separately purchased and/or delivered) operating system(s) - including version number(s).	See Note	1
14-2	Is a list of other third party applications provided by the manufacturer available?	Yes	1
RDMP notes: 1.Vital Images, Inc. does not manufacture hardware or operating system software and /or components which may extend these functions.			
<b>15</b>	<b>SYSTEM AND APPLICATION HARDENING (SAHD)</b>		
	The <b>device's</b> resistance to cyber attacks and <b>malware</b> .		
15-1	Does the <b>device</b> employ any hardening measures? Please indicate in the notes the level of conformance to any industry-recognized hardening standards.	No	—
15-2	Does the <b>device</b> employ any mechanism (e.g., release-specific hash key, checksums, etc.) to ensure the installed program/update is the manufacturer-authorized program or software	Yes	—
15-3	Does the <b>device</b> have external communication capability (e.g., network, modem, etc.)?	Yes	—
15-4	Does the file system allow the implementation of file-level access controls (e.g., New Technology File System (NTFS) for MS Windows platforms)?	Yes	—
15-5	Are all accounts which are not required for the <b>intended use</b> of the <b>device</b> disabled or deleted, for both <b>users</b> and applications?	N/A	—
15-6	Are all shared resources (e.g., file shares) which are not required for the <b>intended use</b> of the <b>device</b> , disabled?	N/A	—
15-7	Are all communication ports which are not required for the <b>intended use</b> of the <b>device</b> closed/disabled?	N/A	1
15-8	Are all services (e.g., telnet, file transfer protocol [FTP], internet information server [IIS], etc.), which are not required for the <b>intended use</b> of the <b>device</b> deleted/disabled? The device is a software package and does not own or control the services or hardware environment on which it is installed	N/A	1
15-9	Are all applications (COTS applications as well as OS-included applications, e.g., MS Internet Explorer, etc.) which are not required for the <b>intended use</b> of the <b>device</b> deleted/disabled?	N/A	—
15-10	Can the <b>device</b> boot from uncontrolled or <b>removable media</b> (i.e., a source other than an internal drive or memory component)?	N/A	—
15-11	Can software or hardware not authorized by the <b>device</b> manufacturer be installed on the device without the use of tools?	N/A	—
SAHD notes: 1.The device is a software package and does not own or control the hardware environment on which it is installed.			
<b>16</b>	<b>SECURITY GUIDANCE (SGUD)</b>		
	The availability of security guidance for <b>operator</b> and administrator of the system and manufacturer sales and service.		
16-1	Are security-related features documented for the <b>device user</b> ?	Yes	—
16-2	Are instructions available for <b>device</b> /media sanitization (i.e., instructions for how to achieve the permanent deletion of personal or other sensitive data)?	N/A	—
SGUD notes:			



© Copyright 2013 by the National Electrical Manufacturers Association and  
the Healthcare Information and Management Systems Society.

HN 1-2013

Device Category	Manufacturer	Document ID	Document Release Date
Clinical image viewer	Vital Images	VLC-10297	Jul-2019
Device Model	Software Revision	Software Release Date	
Vitrea View	7.5	Jul-2019	
Refer to Section 2.3.2 of this standard for the proper interpretation of information requested in this form.			No, N/A, or See
<b>17 HEALTH DATA STORAGE CONFIDENTIALITY (STCF)</b>			
The ability of the <b>device</b> to ensure unauthorized access does not compromise the integrity and confidentiality of <b>private data</b> stored on <b>device</b> or <b>removable media</b> .			
17-1	Can the <b>device</b> encrypt data at rest?	No	—
STCF notes:			
<b>18 TRANSMISSION CONFIDENTIALITY (TXCF)</b>			
The ability of the <b>device</b> to ensure the confidentiality of transmitted <b>private data</b> .			
18-1	Can <b>private data</b> be transmitted only via a point-to-point dedicated cable?	No	—
18-2	Is <b>private data</b> encrypted prior to transmission via a network or <b>removable media</b> ? (If yes, indicate in the notes which encryption standard is implemented.)	No	1
18-3	Is <b>private data</b> transmission restricted to a fixed list of network destinations?	No	—
TXCF notes:			
1.Vitrea View supports HTTPS			
<b>19 TRANSMISSION INTEGRITY (TXIG)</b>			
The ability of the <b>device</b> to ensure the integrity of transmitted <b>private data</b> .			
19-1	Does the <b>device</b> support any mechanism intended to ensure data is not modified during transmission? (If yes, describe in the notes section how this is achieved.)	No	1
TXIG notes:			
1.With HTTPS enabled, this answer is Yes.			
<b>20 OTHER SECURITY CONSIDERATIONS (OTHR)</b>			
Additional security considerations/notes regarding <b>medical device</b> security.			
20-1	Can the <b>device</b> be serviced remotely?	Yes	—
20-2	Can the <b>device</b> restrict remote access to/from specified devices or <b>users</b> or network locations (e.g., specific IP addresses)?	No	—
20-2.1	Can the <b>device</b> be configured to require the local <b>user</b> to accept or initiate remote	No	—

OTHR  
notes:

© Copyright 2013 by the National Electrical Manufacturers Association and  
the Healthcare Information and Management Systems Society.