

## Manufacturer Disclosure Statement for Medical Device Security – MDS<sup>2</sup>

### SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Class II	Vital Images, Inc.	VLC-09204	11-Aug-16
Device Model	Software Revision	Software Release Date	
VitreView	6.11	August, 2016	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	Vital Images, Inc.	5850 Opus Parkway Suite #300, Minnetonka MN 55343	
	Representative Name/Position	952-487-9500	
	Mark Fuchs   Product Manager		

MANAGEMENT OF ELECTRONIC PROTECTED HEALTH INFORMATION (ePHI)	Yes	No	N/A	Note #
1. Can this device transmit or maintain electronic Protected Health Information (ePHI)?	Yes			1
2. Types of ePHI data elements that can be maintained by the device:				
a. Demographic (e.g., name, address, location, unique identification number)?	Yes			
b. Medical record (e.g., medical record #, account #, test or treatment date, device identification number)?	Yes			
c. Diagnostic/therapeutic (e.g., photo/radiograph, test results, or physiologic data with identifying characteristics)?	Yes			
d. Open, unstructured text entered by device user/operator?	No			
3. Maintaining ePHI - Can the device				
a. Maintain ePHI temporarily in volatile memory (i.e., until cleared on by power-off or reset)?	Yes			
b. Store ePHI persistently on local media?	No			
c. Import/export ePHI with other systems?	Yes			
4. Mechanisms used for the transmitting, importing/exporting of ePHI – Can the device				
a. Display ePHI (e.g., video display)?	Yes			
b. Generate hardcopy reports or images containing ePHI?	Yes			
c. Retrieve ePHI from or record ePHI to removable media (e.g., disk, DVD, CD-ROM, tape, CF/SD card, memory stick)?	No			
d. Transmit/receive or import/export ePHI via dedicated cable connection (e.g., IEEE 1073, serial port, USB, FireWire)?	No			
e. Transmit/receive ePHI via a network connection (e.g., LAN, WAN, VPN, intranet, Internet)?	Yes			
f. Transmit/receive ePHI via an integrated wireless connection (e.g. WiFi, Bluetooth, infrared)?	Yes			
g. Other?	N/A			

ADMINISTRATIVE SAFEGUARDS	Yes	No	N/A	Note #
5. Does manufacturer offer operator and technical support training or documentation on device security features?	Yes			2
6. What underlying operating system(s) (including version number) are used by the device?				
Microsoft Windows Server 2008 R2 SP1 Standard x64 Edition				
Microsoft Windows Server 2012 R2 Standard x64 Edition				

PHYSICAL SAFEGUARDS	Yes	No	N/A	Note #
7. Are all device components maintaining ePHI (other than removable media) physically secure (i.e. cannot remove without tools)?			N/A	3
8. Does the device have an integral data backup capability (i.e., backup onto removable media like tape, disk)?			N/A	3
9. Can the device boot from uncontrolled or removable media (i.e., a source other than an internal drive or memory component)?			N/A	3

TECHNICAL SAFEGUARDS	Yes	No	N/A	Note #
10. Can software or hardware not authorized by the device manufacturer be installed on the device without the use of tools?			N/A	3
11. Can the device be serviced remotely (i.e., maintenance activities performed by service person via network or remote connection)?	Yes			4
a. Can the device restrict remote access to specific devices or network locations (e.g., specific IP addresses)?	Yes			3
b. Can the device provide an audit trail of remote-service activity?	No			3
c. Can security patches or other software be installed remotely?	No			3
12. Level of owner/operator service access to device operating system: Can the device owner/operator				
a. Apply device manufacturer-validated security patches?	Yes			5
b. Install or update antivirus software?	Yes			3
c. Update virus definitions on manufacturer-installed antivirus software?	Yes			5
d. Obtain administrative privileges (e.g. access operating system or application via local root or admin account)?	Yes			3
13. Does the device support user/operator specific username and password?	Yes			
14. Does the system force reauthorization after a predetermined length of inactivity (e.g., auto logoff, session lock)?	Yes			3

## Manufacturer Disclosure Statement for Medical Device Security – MDS<sup>2</sup>

### SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Class II	Vital Images, Inc.	VLC-09204	11-Aug-16
Device Model	Software Revision	Software Release Date	
VitreaView	6.11	August, 2016	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	Vital Images, Inc.	5850 Opus Parkway Suite #300, Minnetonka MN 55343	
	Representative Name/Position	952-487-9500	
	Mark Fuchs   Product Manager		

15. Events recorded in device audit trail (e.g., user, date/time, action taken): Can the audit trail record.....		
a. Login and logout by users/operators?.....	Yes	3
b. Viewing of ePHI?.....	Yes	2
c. Creation, modification or deletion of ePHI?.....	No	2
d. Import/export or transmittal/receipt of ePHI?.....	Yes	2
16. Does the device incorporate an emergency access (“break-glass”) feature that is logged?.....	No	3
17. Can the device maintain ePHI during power service interruptions?.....	Yes	3
18. Controls when exchanging ePHI with other devices:.....		
a. Transmitted only via a point-to-point dedicated cable?.....	No	3
b. Encrypted prior to transmission via a network or removable media?.....	Yes	3
c. Restricted to a fixed list of network destinations.....	Yes	3
19. Does the device ensure the integrity of the ePHI data with implicit or explicit error detection/correction technology?.....	No	3

### Other Security Considerations

No additional security concerns identified.

## Manufacturer Disclosure Statement for Medical Device Security – MDS<sup>2</sup>

### SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Class II	Vital Images, Inc.	VLC-09204	11-Aug-16
Device Model	Software Revision	Software Release Date	
VitreaView	6.11	August, 2016	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	Representative Name/Position	5850 Opus Parkway Suite #300, Minnetonka MN 55343	
	Mark Fuchs   Product Manager	952-487-9500	

### SECTION 2

#### **EXPLANATORY NOTES** (from questions 1 - 19)

**IMPORTANT:** Refer to Section 2.2.2 of this standard for the proper interpretation of information requested in this form

1. The term “device” herein refers to the VitreaView<sup>®</sup> software and /or license. Vital Images does not manufacture hardware or operating system software and /or components.
2. The VitreaView<sup>®</sup> software can create audit logs that list user actions to ePHI data. However, VitreaView does not create, modify or delete ePHI.
3. Vital Images, Inc. does not manufacture hardware or operating system software and /or components which may extend these functions.
4. Vital Images, Inc. Technical Support can use secure remote access methods for remote application (VitreaView<sup>®</sup>) support
5. Vital Images, Inc. does not provide customers with Anti-Virus Software Definition Files, Operating System Security patches or Service Packs. Customer is responsible for managing hardware security and software update policies.