



CANON MEDICAL INFORMATICS, INC.

Manufacturer Disclosure Statement for Medical Device Security -- MDS2

Canon Medical Informatics, Inc. Vitrea Advanced Visualization

VLC-11241 Revision A

12-Apr-2024

Question ID	Question	See note
DOC-1	Manufacturer Name	Canon Medical Informatics, Inc. —
DOC-2	Device Description	Visualization Software Device —
DOC-3	Device Model	Vitrea Advanced Visualization —
DOC-4	Document ID	VLC-11241 Revision A —
		5850 Opus Pkwy #300, Minnetonka, MN 55343 —
		(952) 487-9500 —
		Support@mi.medical.canon —
DOC-5	Manufacturer Contact Information	mi.medical.canon —
		Vitrea is a medical diagnostic system that allows the processing, review, analysis, communication, and media interchange of multidimensional digital images acquired from a variety of imaging devices. Vitrea is not meant for primary image interpretation in mammography.
		Vitrea complies with DICOM V3.0 and uses DICOM services for the import of ePHI, study information, and the export of DICOM data, including images and reports, to the desired locations at the facility. All ePHI is stored locally but is not designed for archival purposes. All ePHI data should be removed as desired at the discretion of the hosting facility.
DOC-6	Intended use of device in network-connected environment:	Images are sent and received to and from other DICOM entities (scanners, PACS) via TCP/IP twisted pair Ethernet cable on the network of the facility.
DOC-7	Document Release Date	12-Apr-24
DOC-8	Coordinated Vulnerability Disclosure: Does the manufacturer have a vulnerability disclosure program for this device?	Yes https://global.medical.canon/service-support/securityinformation
DOC-9	ISAO: Is the manufacturer part of an Information Sharing and Analysis Organization?	Yes Parent company Canon is a member of H-ISAC https://h-isac.org/
DOC-10	Diagram: Is a network or data flow diagram available that indicates connections to other system components or expected external resources?	Yes —
DOC-11	SaMD: Is the device Software as a Medical Device (i.e. software-only, no hardware)?	Yes —
DOC-11.1	Does the SaMD contain an operating system?	See Notes Microsoft Windows OS is expected
DOC-11.2	Does the SaMD rely on an owner/operator provided operating system?	See Notes This depends on if hardware is sold along with the medical device. In the case where it is, Windows OS will be provided with the hardware
DOC-11.3	Is the SaMD hosted by the manufacturer?	No
DOC-11.4	Is the SaMD hosted by the customer?	Yes —
		Yes, No, N/A, or See Note
		Note #
	MANAGEMENT OF PERSONALLY IDENTIFIABLE INFORMATION	
MPII-1	Can this device display, transmit, store, or modify personally identifiable information (e.g. electronic Protected Health Information (ePHI))?	Yes —
MPII-2	Does the device maintain personally identifiable information?	Yes
MPII-2.1	Does the device maintain personally identifiable information temporarily in volatile memory (i.e., until cleared by power-off or reset)?	Yes —
MPII-2.2	Does the device store personally identifiable information persistently on internal media?	Yes —
MPII-2.3	Is personally identifiable information preserved in the device's non-volatile memory until explicitly erased?	Yes —
MPII-2.4	Does the device store personally identifiable information in a database?	Yes —

MPII-2.5	Does the device allow configuration to automatically delete local personally identifiable information after it is stored to a long term solution?	Yes	—
MPII-2.6	Does the device import/export personally identifiable information with other systems (e.g., a wearable monitoring device might export personally identifiable information to a server)?	Yes	—
MPII-2.7	Does the device maintain personally identifiable information when powered off, or during power service interruptions?	Yes	—
MPII-2.8	Does the device allow the internal media to be removed by a service technician (e.g., for separate destruction or customer retention)?	Yes	—
MPII-2.9	Does the device allow personally identifiable information records be stored in a separate location from the device's operating system (i.e. secondary internal drive, alternate drive partition, or remote storage location)?	Yes	—
MPII-3	Does the device have mechanisms used for the transmitting, importing/exporting of personally identifiable information?	Yes	—
MPII-3.1	Does the device display personally identifiable information (e.g., video display, etc.)?	Yes	—
MPII-3.2	Does the device generate hardcopy reports or images containing personally identifiable information?	Yes	—
MPII-3.3	Does the device retrieve personally identifiable information from or record personally identifiable information to removable media (e.g., removable-HDD, USB memory, DVD-R/RW, CD-R/RW, tape, CF/SD card, memory stick, etc.)?	Yes	—
MPII-3.4	Does the device transmit/receive or import/export personally identifiable information via dedicated cable connection (e.g., RS-232, RS-423, USB, FireWire, etc.)?	Yes	—
MPII-3.5	Does the device transmit/receive personally identifiable information via a wired network connection (e.g., RJ45, fiber optic, etc.)?	Yes	—
MPII-3.6	Does the device transmit/receive personally identifiable information via a wireless network connection (e.g., WiFi, Bluetooth, NFC, infrared, cellular, etc.)?	No	The customer infrastructure may include wireless networks between the client/server, but no direct WiFi use is performed.
MPII-3.7	Does the device transmit/receive personally identifiable information over an external network (e.g., Internet)?	Yes	—
MPII-3.8	Does the device import personally identifiable information via scanning a document?	No	—
MPII-3.9	Does the device transmit/receive personally identifiable information via a proprietary protocol?	No	—
MPII-3.10	Does the device use any other mechanism to transmit, import or export personally identifiable information?	No	—

Management of Private Data notes:

AUTOMATIC LOGOFF (ALOF)

The device's ability to prevent access and misuse by unauthorized users if device is left idle for a period of time.

ALOF-1	Can the device be configured to force reauthorization of logged-in user(s) after a predetermined length of inactivity (e.g., auto-logoff, session lock, password protected screen saver)?	Yes	—
ALOF-2	Is the length of inactivity time before auto-logoff/screen lock user or administrator	Yes	Configurable range see installation documentation

AUDIT CONTROLS (AUDT)

The ability to reliably audit activity on the device.

AUDT-1	Can the medical device create additional audit logs or reports beyond standard operating system logs?	Yes	—
AUDT-1.1	Does the audit log record a USER ID?	Yes	—

AUDT-1.2	Does other personally identifiable information exist in the audit trail?	Yes	
	Are events recorded in an audit log? If yes, indicate which of the following events are recorded in the audit log:		
AUDT-2	Successful login/logout attempts?	Yes	—
AUDT-2.1	Unsuccessful login/logout attempts?	No	—
AUDT-2.2	Modification of user privileges?	No	—
AUDT-2.3	Creation/modification/deletion of users?	No	—
AUDT-2.4	Presentation of clinical or PII data (e.g. display,	Yes	—
AUDT-2.5	Creation/modification/deletion of data?	Yes	—
AUDT-2.6	Import/export of data from removable media (e.g. USB drive, external hard drive, DVD)?	No	—
AUDT-2.7	Receipt/transmission of data or commands over a network or point-to-point connection?	No	—
AUDT-2.8	Remote or on-site support?	No	—
AUDT-2.8.1	Application Programming Interface (API) and similar activity?	No	—
AUDT-2.8.2	Emergency access?	No	—
AUDT-2.9	Other events (e.g., software updates)?	No	—
AUDT-2.10	Is the audit capability documented in more detail?	No	—
AUDT-2.11	Can the owner/operator define or select which events are recorded in the audit log?	No	—
AUDT-3	Is a list of data attributes that are captured in the audit log for an event available?	No	—
AUDT-4	Does the audit log record date/time?	Yes	—
AUDT-4.1	Can date and time be synchronized by Network Time Protocol (NTP) or equivalent time source?	Yes	—
AUDT-4.1.1	Can audit log content be exported?	Yes	—
AUDT-5	Via physical media?	Yes	—
AUDT-5.1	Via IHE Audit Trail and Node Authentication (ATNA) profile to SIEM?	No	—
AUDT-5.2	Via Other communications (e.g., external service device, mobile applications)?	No	—
AUDT-5.3	Are audit logs encrypted in transit or on storage media?	No	—
AUDT-5.4	Can audit logs be monitored/reviewed by owner/operator?	Yes	—
AUDT-6	Are audit logs protected from modification?	No	—
AUDT-7	Are audit logs protected from access?	No	—
AUDT-7.1	Can audit logs be analyzed by the device?	No	—
AUDT-8			

AUTHORIZATION (AUTH)

The ability of the device to determine the authorization of users.

AUTH-1	Does the device prevent access to unauthorized users through user login requirements or other mechanism?	Yes	Technical safeguards such as password protected account and group membership
	Can the device be configured to use federated credentials management of users for authorization (e.g., LDAP, OAuth)?	Yes	Microsoft Active Directory LDAP
AUTH-1.1	Can the customer push group policies to the device (e.g., Active Directory)?	Yes	—
AUTH-1.2	Are any special groups, organizational units, or group policies required?	Yes	—
AUTH-1.3	Can users be assigned different privilege levels based on 'role' (e.g., user, administrator, and/or service, etc.)?	Yes	—
AUTH-2	Can the device owner/operator grant themselves unrestricted administrative privileges (e.g., access operating system or application via local root or administrator account)?	Yes	—
AUTH-3	Does the device authorize or control all API access requests?	Yes	—
AUTH-4	Does the device run in a restricted access mode, or 'kiosk mode', by default?	Yes	—
AUTH-5			

CYBER SECURITY PRODUCT UPGRADES (CSUP)

The ability of on-site service staff, remote service staff, or authorized customer staff to install/upgrade device's security patches.

CSUP-1	Does the device contain any software or firmware which may require security updates during its operational life, either from the device manufacturer or from a third-party manufacturer of the software/firmware? If no, answer "N/A" to questions in this section.	Yes	—
CSUP-2	Does the device contain an Operating System? If yes, complete 2.1-2.4.	Yes	Windows OS is provided if hardware is included in medical device sale
CSUP-2.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	Yes	—
CSUP-2.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	Microsoft Windows updates can be applied as they are released. All others should be vetted through vendor.
CSUP-2.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—
CSUP-2.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	—
CSUP-3	Does the device contain Drivers and Firmware? If yes, complete 3.1-3.4.	Yes	NVidia
CSUP-3.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	Yes	—
CSUP-3.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	No	—
CSUP-3.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—
CSUP-3.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	—
CSUP-4	Does the device contain Anti-Malware Software? If yes, complete 4.1-4.4.	Yes	Windows Defender
CSUP-4.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	Yes	—
CSUP-4.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	No	These come through Windows updates
CSUP-4.3	Does the device have the capability to receive remote installation of patches or software updates?	Yes	Yes via Windows Updates
CSUP-4.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	—
CSUP-5	Does the device contain Non-Operating System commercial off-the-shelf components? If yes, complete 5.1-5.4.	Yes	—
CSUP-5.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	See Notes	For off-the-shelf components that are allowed to be updated, instructions are included in Administration Guides
CSUP-5.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	—
CSUP-5.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—
CSUP-5.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	Yes	—
CSUP-6	Does the device contain other software components (e.g., asset management software, license management)? If yes, please provide details or reference in notes and complete 6.1-6.4.	See Notes	Remote Desktop and SQL Licensing is the customer's responsibility The device includes licensing management for the device
CSUP-6.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	N/A	—
CSUP-6.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	N/A	—

CSUP-6.3	Does the device have the capability to receive remote installation of patches or software updates?	N/A	—
CSUP-6.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	N/A	—
CSUP-7	Does the manufacturer notify the customer when updates are approved for installation?	No	—
CSUP-8	Does the device perform automatic installation of software updates?	No	—
CSUP-9	Does the manufacturer have an approved list of third-party software that can be installed on the device?	No	—
CSUP-10	Can the owner/operator install manufacturer-approved third-party software on the device themselves?	Yes	—
CSUP-10.1	Does the system have mechanism in place to prevent installation of unapproved software?	No	—
CSUP-11	Does the manufacturer have a process in place to assess device vulnerabilities and updates?	Yes	—
CSUP-11.1	Does the manufacturer provide customers with review and approval status of updates?	No	—
CSUP-11.2	Is there an update review cycle for the device?	Yes	Microsoft Windows updates can be applied as they are released.

HEALTH DATA DE-IDENTIFICATION (DIDT)

The ability of the device to directly remove information that allows identification of a person.

DIDT-1	Does the device provide an integral capability to de-identify personally identifiable information?	Yes	—
DIDT-1.1	Does the device support de-identification profiles that comply with the DICOM standard for de-identification?	Yes	—

DATA BACKUP AND DISASTER RECOVERY (DTBK)

The ability to recover after damage or destruction of device data, hardware, software, or site configuration information.

DTBK-1	Does the device maintain long term primary storage of personally identifiable information / patient information (e.g. PACS)?	No	—
DTBK-2	Does the device have a “factory reset” function to restore the original device settings as provided by the manufacturer?	No	—
DTBK-3	Does the device have an integral data backup capability to removable media?	Yes	DVD, BluRay
DTBK-4	Does the device have an integral data backup capability to remote storage?	No	—
DTBK-5	Does the device have a backup capability for system configuration information, patch restoration, and software restoration?	No	—
DTBK-6	Does the device provide the capability to check the integrity and authenticity of a backup?	No	—

EMERGENCY ACCESS (EMRG)

The ability of the device user to access personally identifiable information in case of a medical emergency situation that requires immediate access to stored personally identifiable information.

EMRG-1	Does the device incorporate an emergency access (i.e. “break-glass”) feature?	No	—
--------	---	----	---

HEALTH DATA INTEGRITY AND AUTHENTICITY (IGAU)

How the device ensures that the stored data on the device has not been altered or destroyed in a non-authorized manner and is from the originator.

IGAU-1	Does the device provide data integrity checking mechanisms of stored health data (e.g., hash or digital signature)?	No	—
IGAU-2	Does the device provide error/failure protection and recovery mechanisms for stored health data (e.g., RAID-5)?	Yes	Raid 10

MALWARE DETECTION/PROTECTION (MLDP)

The ability of the device to effectively prevent, detect and remove malicious software (malware).

MLDP-1	Is the device capable of hosting executable software? Does the device support the use of anti-malware software (or other anti-malware mechanism)?	Yes	—
MLDP-2	Provide details or reference in notes. Does the device include anti-malware software by default?	Yes	—
MLDP-2.1	Does the device have anti-malware software available as an option?	Yes	Delivered with Windows OS Windows Defender
MLDP-2.2	Does the device documentation allow the owner/operator to install or update anti-malware software?	No	—
MLDP-2.3	Can the device owner/operator independently (re-)configure anti-malware settings?	Yes	—
MLDP-2.4	Does notification of malware detection occur in the device user interface?	Yes	—
MLDP-2.5	Can only manufacturer-authorized persons repair systems when malware has been detected?	Yes	Windows Defaults are used. Can be configured by customer.
MLDP-2.6	Are malware notifications written to a log?	No	—
MLDP-2.7	Are there any restrictions on anti-malware (e.g., purchase, installation, configuration, scheduling)?	Yes	See Microsoft documentation
MLDP-2.8	If the answer to MLDP-2 is NO, and anti-malware cannot be installed on the device, are other compensating controls in place or available?	No	—
MLDP-3	Does the device employ application whitelisting that restricts the software and services that are permitted to be run on the device?	N/A	—
MLDP-4	Does the device employ a host-based intrusion detection/prevention system?	No	—
MLDP-5	Can the host-based intrusion detection/prevention system be configured by the customer?	No	—
MLDP-5.1	Can a host-based intrusion detection/prevention system be installed by the customer?	N/A	—
MLDP-5.2		Yes	—

NODE AUTHENTICATION (NAUT)

The ability of the device to authenticate communication partners/nodes.

NAUT-1	Does the device provide/support any means of node authentication that assures both the sender and the recipient of data are known to each other and are authorized to receive transferred information (e.g. Web APIs, SMTP, SNMP)?	Yes	Web APIs between Advanced Servers and Management node for Enterprise Deployments. N/A for other deployment types.
NAUT-2	Are network access control mechanisms supported (E.g., does the device have an internal firewall, or use a network connection white list)?	Yes	—
NAUT-2.1	Is the firewall ruleset documented and available for review?	Yes	—
NAUT-3	Does the device use certificate-based network connection authentication?	Yes	—

CONNECTIVITY CAPABILITIES (CONN)

All network and removable media connections must be considered in determining appropriate security controls. This section lists connectivity capabilities that may be present on the device.

CONN-1	Does the device have hardware connectivity capabilities?	Yes	Windows based workstations and servers using TCP/IP protocols for communication
CONN-1.1	Does the device support wireless connections?	No	—
CONN-1.1.1	Does the device support Wi-Fi?	N/A	—
CONN-1.1.2	Does the device support Bluetooth?	N/A	—
CONN-1.1.3	Does the device support other wireless network connectivity (e.g. LTE, Zigbee, proprietary)?	N/A	—
CONN-1.1.4	Does the device support other wireless connections (e.g., custom RF controls, wireless detectors)?	N/A	—
CONN-1.2	Does the device support physical connections?	Yes	—
CONN-1.2.1	Does the device have available RJ45 Ethernet ports?	Yes	—
CONN-1.2.2	Does the device have available USB ports?	Yes	System can come with or without USB, USB can be disabled, not required.
CONN-1.2.3	Does the device require, use, or support removable memory devices?	Yes	System can use removable devices but not required.
CONN-1.2.4	Does the device support other physical connectivity?	No	—
CONN-2	Does the manufacturer provide a list of network ports and protocols that are used or may be used on the device?	Yes	See installation documentation
CONN-3	Can the device communicate with other systems within the customer environment?	Yes	—
CONN-4	Can the device communicate with other systems external to the customer environment (e.g., a service host)?	No	— Vitrea Enterprise deployments do make and can receive API calls. Extend and Workstation deployments do not.
CONN-5	Does the device make or receive API calls?	Yes	
CONN-6	Does the device require an internet connection for its intended use?	No	—
CONN-7	Does the device support Transport Layer Security (TLS)?	Yes	—
CONN-7.1	Is TLS configurable?	Yes	
CONN-8	Does the device provide operator control functionality from a separate device (e.g., telemedicine)?	Yes	

PERSON AUTHENTICATION (PAUT)

The ability to configure the device to authenticate users.

PAUT-1	Does the device support and enforce unique IDs and passwords for all users and roles (including service accounts)?	Yes	—
PAUT-1.1	Does the device enforce authentication of unique IDs and passwords for all users and roles (including service accounts)?	Yes	—
PAUT-2	Is the device configurable to authenticate users through an external authentication service (e.g., MS Active Directory, NDS, LDAP, OAuth, etc.)?	Yes	—
PAUT-3	Is the device configurable to lock out a user after a certain number of unsuccessful logon attempts?	Yes	Determined by customer deployment
PAUT-4	Are all default accounts (e.g., technician service accounts, administrator accounts) listed in the documentation?	Yes	—
PAUT-5	Can all passwords be changed?	Yes	—
PAUT-6	Is the device configurable to enforce creation of user account passwords that meet established (organization specific) complexity rules?	Yes	—
PAUT-7	Does the device support account passwords that expire periodically?	Yes	—
PAUT-8	Does the device support multi-factor authentication?	No	—
PAUT-9	Does the device support single sign-on (SSO)?	Yes	—
PAUT-10	Can user accounts be disabled/locked on the device?	Yes	—
PAUT-11	Does the device support biometric controls?	No	—
PAUT-12	Does the device support physical tokens (e.g. badge access)?	Yes	SmartCard
PAUT-13	Does the device support group authentication (e.g. hospital teams)?	Yes	Via MS Active Directory or Windows auth of user groups
PAUT-14	Does the application or device store or manage authentication credentials?	No	Workstation, Extend, and ESS deployments not on a domain delegate credentials to Windows
PAUT-14.1	Are credentials stored using a secure method?	N/A	—

PHYSICAL LOCKS (PLOK)

Physical locks can prevent unauthorized users with physical access to the device from compromising the integrity and confidentiality of personally identifiable information stored on the device or on removable media

PLOK-1	Is the device software only? If yes, answer "N/A" to remaining questions in this section.	Yes	—
PLOK-2	Are all device components maintaining personally identifiable information (other than removable media) physically secure (i.e., cannot remove without tools)?	N/A	Physically securing the hardware is the customers responsibility.
PLOK-3	Are all device components maintaining personally identifiable information (other than removable media) physically secured behind an individually keyed locking device?	N/A	Physically securing the hardware is the customers responsibility.
PLOK-4	Does the device have an option for the customer to attach a physical lock to restrict access to removable media?	N/A	Physically securing the hardware is the customers responsibility.

ROADMAP FOR THIRD PARTY COMPONENTS IN DEVICE LIFE CYCLE (RDMP)

Manufacturer's plans for security support of third-party components within the device's life cycle.

RDMP-1	Was a secure software development process, such as ISO/IEC 27034 or IEC 62304, followed during product development?	Yes	IEC 62304
RDMP-2	Does the manufacturer evaluate third-party applications and software components included in the device for secure development practices?	Yes	—
RDMP-3	Does the manufacturer maintain a web page or other source of information on software support dates and updates?	Yes	—
RDMP-4	Does the manufacturer have a plan for managing third-party component end-of-life?	Yes	—

SOFTWARE BILL OF MATERIALS (SBoM)

A Software Bill of Material (SBoM) lists all the software components that are incorporated into the device being described for the purpose of operational security planning by the healthcare delivery organization. This section supports controls in the RDMP section.

SBOM-1	Is the SBoM for this product available?	Yes	—
SBOM-2	Does the SBoM follow a standard or common method in describing software components?	Yes	CycloneDX format
SBOM-2.1	Are the software components identified?	Yes	—
SBOM-2.2	Are the developers/manufacturers of the software components identified?	Yes	—
SBOM-2.3	Are the major version numbers of the software components identified?	Yes	—
SBOM-2.4	Are any additional descriptive elements identified?	No	—
SBOM-3	Does the device include a command or process method available to generate a list of software components installed on the device?	No	—
SBOM-4	Is there an update process for the SBoM?	Yes	—

SYSTEM AND APPLICATION HARDENING (SAHD)

The device's inherent resistance to cyber attacks and malware.

SAHD-1	Is the device hardened in accordance with any industry standards?	Yes	NIST 800-53
SAHD-2	Has the device received any cybersecurity certifications?	No	—
SAHD-3	Does the device employ any mechanisms for software integrity checking?	Yes	MD5 Checksum on download

SAHD-3.1	Does the device employ any mechanism (e.g., release-specific hash key, checksums, digital signature, etc.) to ensure the installed software is manufacturer-authorized?	Yes	—
SAHD-3.2	Does the device employ any mechanism (e.g., release-specific hash key, checksums, digital signature, etc.) to ensure the software updates are the manufacturer-authorized updates?	Yes	—
SAHD-4	Can the owner/operator perform software integrity checks (i.e., verify that the system has not been modified or tampered with)?	No	—
SAHD-5	Is the system configurable to allow the implementation of file-level, patient level, or other types of access controls?	No	—
SAHD-5.1	Does the device provide role-based access controls?	Yes	—
SAHD-6	Are any system or user accounts restricted or disabled by the manufacturer at system delivery?	No	—
SAHD-6.1	Are any system or user accounts configurable by the end user after initial configuration?	Yes	—
SAHD-6.2	Does this include restricting certain system or user accounts, such as service technicians, to least privileged access?	Yes	—
SAHD-7	Are all shared resources (e.g., file shares) which are not required for the intended use of the device disabled?	Yes	—
SAHD-8	Are all communication ports and protocols that are not required for the intended use of the device disabled?	Yes	Yes for specific deployments including Extend.
SAHD-9	Are all services (e.g., telnet, file transfer protocol [FTP], internet information server [IIS], etc.), which are not required for the intended use of the device deleted/disabled?	Yes	—
SAHD-10	Are all applications (COTS applications as well as OS-included applications, e.g., MS Internet Explorer, etc.) which are not required for the intended use of the device deleted/disabled?	No	—
SAHD-11	Can the device prohibit boot from uncontrolled or removable media (i.e., a source other than an internal drive or memory component)?	Yes	—
SAHD-12	Can unauthorized software or hardware be installed on the device without the use of physical tools?	Yes	—
SAHD-13	Does the product documentation include information on operational network security scanning by users?	No	—
SAHD-14	Can the device be hardened beyond the default provided state?	Yes	—
SAHD-14.1	Are instructions available from vendor for increased hardening?	Yes	—
SHAD-15	Can the system prevent access to BIOS or other bootloaders during boot?	Yes	—
SAHD-16	Have additional hardening methods not included in 2.3.19 been used to harden the device?	No	—

SECURITY GUIDANCE (SGUD)

Availability of security guidance for operator and administrator of the device and manufacturer sales and service.

SGUD-1	Does the device include security documentation for the owner/operator?	Yes	Installation and User Documentation included with the product
SGUD-2	Does the device have the capability, and provide instructions, for the permanent deletion of data from the device or media?	See Notes	It has the capability but no specific directions are provided.
SGUD-3	Are all access accounts documented?	Yes	—
SGUD-3.1	Can the owner/operator manage password control for all accounts?	Yes	—
SGUD-4	Does the product include documentation on recommended compensating controls for the device?	Yes	—

HEALTH DATA STORAGE CONFIDENTIALITY (STCF)

The ability of the device to ensure unauthorized access does not compromise the integrity and confidentiality of personally identifiable information stored on the device or removable media.

STCF-1	Can the device encrypt data at rest?	Yes	—
STCF-1.1	Is all data encrypted or otherwise protected?	No	
STCF-1.2	Is the data encryption capability configured by default?	No	
STCF-1.3	Are instructions available to the customer to configure encryption?	Yes	
STCF-2	Can the encryption keys be changed or configured?	Yes	—
STCF-3	Is the data stored in a database located on the device?	See Notes	This is deployment dependent. The Vitrea Enterprise deployment can be onbox or offbox at customer request. All other deployments the database on the device.
STCF-4	Is the data stored in a database external to the device?	See Notes	This is deployment dependent. The Vitrea Enterprise deployment can be onbox or offbox at customer request. All other deployments the database on the device.

TRANSMISSION CONFIDENTIALITY (TXCF)

The ability of the device to ensure the confidentiality of transmitted personally identifiable information.

TXCF-1	Can personally identifiable information be transmitted only via a point-to-point dedicated cable?	No	—
TXCF-2	Is personally identifiable information encrypted prior to transmission via a network or removable media?	No	—
TXCF-2.1	If data is not encrypted by default, can the customer configure encryption options?	Yes	—
TXCF-3	Is personally identifiable information transmission restricted to a fixed list of network destinations?	Yes	DICOM destinations are configured by admin only
TXCF-4	Are connections limited to authenticated systems?	Yes	—
TXCF-5	Are secure transmission methods supported/implemented (DICOM, HL7, IEEE 11073)?	No	—

TRANSMISSION INTEGRITY (TXIG)

The ability of the device to ensure the integrity of transmitted data.

TXIG-1	Does the device support any mechanism (e.g., digital signatures) intended to ensure data is not modified during transmission?	No	—
TXIG-2	Does the device include multiple sub-components connected by external cables?	No	—

REMOTE SERVICE (RMOT)

Remote service refers to all kinds of device maintenance activities performed by a service person via network or other remote connection.

RMOT-1	Does the device permit remote service connections for device analysis or repair?	Yes	—
RMOT-1.1	Does the device allow the owner/operator to initiate remote service sessions for device analysis or repair?	Yes	—
RMOT-1.2	Is there an indicator for an enabled and active remote session?	Yes	—
RMOT-1.3	Can patient data be accessed or viewed from the device during the remote session?	Yes	—
RMOT-2	Does the device permit or use remote service connections for predictive maintenance data?	See Notes	Solution Health software implementation can be used with the Vitrea software for predictive maintenance.
RMOT-3	Does the device have any other remotely accessible functionality (e.g. software updates, remote	No	—

OTHER SECURITY CONSIDERATIONS (OTHR)

NONE

Notes:

Example note. Please keep individual notes to one cell. Please use separate notes for separate information

Note 1